

Month/Day/Year

URGENT FIELD SAFETY NOTICE
Datascope Intra-Aortic Balloon Pumps (IABP)
Battery Usage, Charging, Maintenance and Storage Instructions

AFFECTED PRODUCT	PART NUMBER	DISTRIBUTION DATE
Cardiosave Hybrid IABP Cardiosave Rescue IABP CS300 IABP CS100 IABP	All	All

PLEASE FORWARD THIS INFORMATION TO ALL CURRENT AND POTENTIAL CARDIOSAVE HYBRID, CARDIOSAVE RESCUE, CS300 and CS100 INTRA-AORTIC BALLOON PUMP (IABP) USERS WITHIN YOUR FACILITY.

IF YOU ARE A DISTRIBUTOR WHO HAS SHIPPED ANY AFFECTED PRODUCTS TO CUSTOMERS, PLEASE FORWARD THIS DOCUMENT TO THEIR ATTENTION FOR APPROPRIATE ACTION.

Dear Customer,

This is to notify you that the Datascope IABP devices(s) (Cardiosave Hybrid IABP, Cardiosave Rescue IABP, CS300 IABP and CS100 IABP) your facility may have received from Getinge are part of a field correction initiated May 16, 2019. This field correction is being conducted to ensure that all IABP users follow each device's Operating Instructions Manual for recommendations on usage, charging, maintenance and storage of the batteries, as battery run times and discharge cycles vary between IABP models. If battery maintenance is not performed per the Operating Instructions Manual for each IABP, the battery may provide less than the expected minimum run time of operating power per battery. In addition, Getinge will proactively reach out to all customers to help ensure this field correction is adequate and effective to reduce IABP shutdown while operating on battery power.

There have been five patient deaths reported since March 2016, although the deaths cannot be definitively attributed to the device shutting down while operating on battery power.

There is patient risk of hemodynamic instability due to sudden interruption or temporary suspension of therapy. In patients with mild to moderate hemodynamic compromise, inotropic agents can provide sufficient hemodynamic support while the unit is reconnected to an AC source or alternative therapy is initiated. Therefore, an interruption of the therapy would be unlikely to lead to a life-threatening situation. However, in critical patients with severely compromised hemodynamic function dependent on continuous circulatory support, an interruption or delay in IAB support as a result of an unexpected shutdown or failure to initiate therapy can occasionally/likely have more severe consequences that can be life threatening.

Immediate Interim actions to be taken by User:

- Ensure the IABP is plugged into an AC power outlet whenever possible during patient use to prevent the battery from depleting.
- Ensure the IABP is plugged into an AC power outlet when the system is not in use. The batteries should be kept at a full charge even when the IABP is not in use.
- When transporting patients within or between facilities, please refer to the IABP Operating Instructions Manual for recommendations on portable/battery operation. For example:

- Prior to portable operation, the battery should be fully charged
- For Cardiosave Rescue and Cardiosave Hybrid only:
 - Additional charged batteries should be on hand during transport
 - Ensure the batteries are properly seated in the battery compartment/charger and the IABP Console is completely seated/secured into the IABP Cart
 - For Cardiosave Hybrid, you can verify if the Console is completely seated in the IABP cart by the indicator on the display:



If the word “**Hybrid**” is displayed, then the IABP console is secured into the IABP cart.

If the word “**Rescue**” is displayed, then the IABP console is not secured into the IABP cart

- Check battery run time and replace batteries as required, as recommended in each IABPs Operating Instructions Manual. A reduction in run time can occur over a battery's life for reasons such as age, storage temperature and discharge depth. Batteries should be replaced:
 - After reaching the maximum number of charge-discharge cycles
 - When the battery provides less than the minimum specified run time
 - If the battery is broken, cracked, leaking or damaged
 - When the labeled lifetime of the battery is reached

NOTE: Batteries for the Cardiosave Hybrid and Cardiosave Rescue IABPs sold before [INSERT MONTH/YEAR REPRESENTING 4 YEARS BEFORE THE MONTH/YEAR OF THE LETTER DATE] should be replaced immediately as the labeled lifetime for these batteries is 4 years. Replacement batteries can be ordered through your sales or service representative. To determine the date of manufacture for all Cardiosave batteries, refer to attached document, ‘Cardiosave Lithium-ion Battery Pack’ ML-0795.

NOTE: CS100/CS300: Informational messages on the display screen provide information to the operator regarding the batteries. The **Battery Maintenance Required** message indicates that the IABP internal battery requires maintenance. The Battery test due date or Battery Replacement Date predate the current system date at startup or the internal battery has a total accumulated discharge time in excess of 100 total discharge cycles.

For all replacement batteries, ensure only Datascope approved/sourced batteries are installed/used.

In case of a sudden shutdown of an IABP such as battery depletion is related to the static condition (no inflating or deflating) of the balloon during the interruption of therapy, it is important to note the following WARNING in the Operating Instructions for all Datascope IABPs:

WARNING: *The patient balloon should not remain inactive in the patient (i.e., no inflating or deflating) for more than 30 minutes, due to the potential for thrombus formation.*

In the unlikely event that this situation was to occur, transfer the patient to an alternative Datascope IABP. If an alternative Datascope IABP is unavailable; manually inflate the IAB with air or helium and immediately aspirate. Please refer to the IAB Instructions for Use, Manually Inflating and Deflating a Catheter. The IAB Instructions for Use reiterates that a catheter should not remain inactive for more than 30 minutes, due to the potential for thrombus formation. Alternatively, the IAB could be removed.

To support our customers in this field correction, Getinge will execute the following Corrective Actions:

Initial Action - Cardiosave Hybrid IABP, Cardiosave Rescue IABP, CS300 IABP and CS100 IABP:

- Getinge has developed a battery operations, care and maintenance reference guide specific to the IABP(s) based on the Operating Instructions Manual(s) provided with each device. These guides are available by accessing the link provided below:

info.getinge.com/ca-batteryguides

A hard copy of the guides are available upon request by contacting your local Sales and Service Representative.

Additional Actions - Cardiosave Hybrid IABP, Cardiosave Rescue IABP:

- Getinge is currently developing a Cardiosave battery maintenance software upgrade targeted for early 2020. This updated software requires FDA clearance and once completed, a Getinge Service representative will contact you to schedule the installation of the updated software. This work will be done at no cost to your facility.
- NOTE: A similar software upgrade was released for the CS300 IABP and CS100 IABP in 2017. If you are unsure whether your IABP has been updated with the released software upgrade, please contact your Getinge Sales & Service Office with the Model and Serial number of the IABP. The Sales & Service Office will determine if the IABP software has been updated.

Please complete the attached Urgent Field Safety Notice Response Form (page 4) to acknowledge that you have received this Field Safety Notice. Please fax the completed form to **INSERT LOCAL SSU FAX NUMBER** or send via email to **INSERT LOCAL SSU EMAIL ADDRESS**.

Maquet/Datascope apologizes for any inconvenience you may experience as a result of this Medical Device Correction. If you have any questions, please contact your local Maquet/Getinge representative.

Thank you for your cooperation and immediate assistance.

Sincerely,

[NAME]
[TITLE]

Month/Day/Year

URGENT FIELD SAFETY NOTICE
RESPONSE FORM
Datascope Intra-Aortic Balloon Pumps (IABP)
Battery Usage, Charging, Maintenance and Storage Instructions

Please return the completed form by email to INSERT LOCAL SSU EMAIL ADDRESS or
by FAX to INSERT LOCAL SSU FAX NUMBER

<u>AFFECTED PRODUCT</u>	<u>PART NUMBER</u>	<u>DISTRIBUTION DATE</u>
Cardiosave Hybrid IABP Cardiosave Rescue IABP CS300 IABP CS100 IABP	All	All

[ACCOUNT NO.]
[ENTER FACILITY NAME]
[ADDRESS]

I acknowledge that I have reviewed and understand this Urgent Field Safety Notice for the affected Datascope Intra-Aortic Balloon Pump(s) at this facility.

I confirm that all users of the Datascope Intra-Aortic Balloon Pump(s) at this facility have been notified accordingly.

Facility Representative:

Signature: _____ Date: _____

Name: _____ Phone: _____

Title: _____ Department: _____

Facility Name: _____

Address, City and State: _____

EMAIL to INSERT LOCAL SSU EMAIL ADDRESS or FAX to INSERT LOCAL SSU FAX NUMBER.