

Urgent Field Safety Notice

ACHC20-06.A.OUS.DM

May 2020

Dimension® clinical chemistry systems

N-acetyl-p-benzoquinone imine (NAPQI) Interference with Dimension® Enzymatic Creatinine (EZCR) Flex® reagent cartridge

Our records indicate that your facility may have received the following product:

Table 1. Dimension® affected product

Assay	Test Code	Catalog Number	Siemens Material Number (SMN)	Lot Number
Enzymatic Creatinine	EZCR	DF270B	10471520	ALL

Reason for Correction

The purpose of this communication is to inform you of an interference identified with the product indicated in Table 1 above and to provide instructions on actions that your laboratory must take.

Siemens Healthcare Diagnostics has become aware that the presence of N-acetyl-p-benzoquinone imine (NAPQI), a metabolite of acetaminophen, may cause interference in certain chemistry assays when testing patient samples. Through spiking studies of NAPQI at 15 mg/L [101 µmol/L], Siemens observed a bias with Dimension EZCR of approximately -12% at a creatinine concentration of 1.02 mg/dL (90 µmol/L). This level of NAPQI correlates to a toxic level of acetaminophen. Titration experiments were subsequently performed to characterize the potential for interference at decreasing concentrations of NAPQI.

Actions to be Taken by Siemens:

The "Limitations of the Procedure" section of the Dimension EZCR assay Instructions for Use (IFU) will be updated to indicate that *'N-acetyl-p-benzoquinone imine (NAPQI) is a metabolite of acetaminophen. NAPQI concentrations of approximately 10 mg/L correlating to toxic levels of acetaminophen demonstrates a ≤10% change in results. NAPQI concentrations greater than this may lead to falsely depressed results for patient samples.'*

The information related to NAPQI provided in this letter supersedes the information in the current Dimension EZCR IFU until it is updated.

The updated IFUs will be uploaded into Document Library where all registered users who opt in to receive alerts will be notified of the updated IFU.

N-acetyl-p-benzoquinone imine (NAPQI) Interference with Dimension® Enzymatic Creatinine (EZCR)

Risk to Health

There is negligible risk to health. At toxic levels of acetaminophen, NAPQI interference may lead to falsely depressed creatinine values, however the magnitude of the bias would not be expected to lead to a clinically significant difference in patient management.

Actions to be Taken by the Customer:

- Be aware of the limitation indicated above.
- Please review this letter with your Medical Director.
- Complete and return the Field Correction Effectiveness Check Form attached to this letter within 30 days.
- If you have received any complaints of illness or adverse events associated with the products listed in Table 1, immediately contact your local Siemens Customer Care Center or your local Siemens technical support representative.

Please retain this letter with your laboratory records and forward this letter to those who may have received this product.

We apologize for the inconvenience this situation may cause. If you have any questions, please contact your Siemens Customer Care Center or your local Siemens Technical Support representative.

Dimension is a trademark of Siemens Healthcare Diagnostics.

N-acetyl-p-benzoquinone imine (NAPQI) Interference with Dimension® Enzymatic Creatinine (EZCR)

Frequently Asked Questions:

1. Which other assays have been evaluated for NAPQI interference?

The assays indicated below were tested on the Siemens ADVIA Chemistry and Atellica CH systems. These assays have identical principles of procedure to the equivalent assays on the Dimension® clinical chemistry system. No significant interference is expected on the following Dimension assays: Cholesterol, Creatinine (Jaffe), Direct HDL, Lactic Acid, Lipase, Triglyceride, and Uric Acid.

Additionally, Direct and Total Bilirubin (DBI and TBI) assays were tested on the Dimension system and showed show no interference ($\leq 10\%$ bias) for NAPQI concentrations of >15 mg/L ($101 \mu\text{mol/L}$) at sample concentrations similar to those published in each product's IFU.

2. Is the Jaffe Creatinine assay impacted in the presence of NAPQI?

Siemens testing has shown that the Jaffe methodology is not impacted by NAPQI interference as the Jaffe methodology uses different reagents and parameters than the EZCR assay.

N-acetyl-p-benzoquinone imine (NAPQI) Interference with Dimension® Enzymatic Creatinine (EZCR)

FIELD CORRECTION EFFECTIVENESS CHECK

N-acetyl-p-benzoquinone imine (NAPQI) Interference with Dimension® Enzymatic Creatinine (EZCR) Flex® reagent cartridge

This response form is to confirm receipt of the enclosed Siemens Healthcare Diagnostics Urgent Field Safety Notice ACHC20-06.A.OUS.DM dated May 2020 regarding NAPQI Interference with Dimension® Enzymatic Creatinine (EZCR). Please read the question and indicate the appropriate answer.

Return this completed form to Siemens Healthcare Diagnostics as per the instructions provided at the bottom of this page.

1. I have read and understood the Urgent Field Safety Notice provided in this letter. Yes No

Name of person completing questionnaire:

Date:

Title:

Institution: Instrument Serial Number:

Street:

City: State:

Phone: Country:

Please send a scanned copy of the completed form via email to XXXX@XXXX

Or to fax this completed form to the Customer Care Center at XXXXXX

If you have any questions, contact your local Siemens Healthineers technical support representative.