



**Subsidiary/Distributor Actions:****Inform end-users by carrying out the following actions:**

1. Translate the customer advisory letters into your local language(s) and print it on your official company paper.
2. Compose complete list of affected customers including serial numbers of analyzers. For subsidiaries and distributors covering more countries, the list must be divided into countries.
3. Contact each affected customer as follows,
  - Submit the customer advisory letter to the distributors and customers, or
  - Visit the customers to hand over the letter and explain the problem.
4. Consolidate status of receipt of Recall Response Forms from affected customers in "ABL800 Response and Upgrade Sheet" (excel) enclosed with this FAN.
5. Schedule a visit to each affected customer and upgrade the analyzer software to version 6.19 or higher.
6. Consolidate upgrade status for all affected analyzers in "ABL800 Response and Upgrade Sheet" (excel) enclosed with this FAN, and submit to RMED.

**Please carry out the following action for new customers:**

Ensure that software version 6.19 (or higher) is installed prior to delivery.

**Completion Dates:** The actions must be completed by the dates stated:

- Actions 1 and 2 must be completed and confirmed to RMED before **March 6, 2020** (by submitting the translated customer advisory letter and the list of affected customers).
- Action 3 must be completed and confirmed to RMED before **April 8, 2020**. (by returning the FAC1 - Customer advisory letter).
- Action 4 must be completed and confirmed to RMED before **May 7, 2020** (by returning FAC2 – Customer response and the partially completed "ABL800 Response and Upgrade Sheet" (excel).
- Actions 5 and 6 must be completed and confirmed to RMED before **February 20, 2021** (by returning the FAC3 – Upgrade and by submitting completed "ABL800 Response and Upgrade Sheet" to RMED).

**Tools:****The following tools are available:**

- Customer advisory letter
- Software version 6.19, 933-784  
Available for download in Global Services Library
- ABL800 Response and Upgrade Excel sheet

**Inquiries:** Please refer to the below departments for inquiries related to this Field Action Note:

- For technical, commercial, and practical questions please contact RMED Technical Product Support and Service:  
Email: [technical.support@radiometer.dk](mailto:technical.support@radiometer.dk) or  
Telephone: +45 4010 8827
- For questions from your local national competent authorities please contact RMED Vigilance:  
Email: [vigilance@radiometer.dk](mailto:vigilance@radiometer.dk)
- To confirm receipt and submit customer lists, translated letters, FACs, Customer Response Sheets, etc., please use:  
Email: [fan@radiometer.dk](mailto:fan@radiometer.dk)

---

<b>Regulatory:</b>	For regulatory reasons all subsidiaries and distributors must email complete customer lists for each country to RMED.
<b>USA/Canada:</b>	The affected product, ABL800, has been distributed to USA and Canada and the field action will be reported to the FDA and Health Canada.
<b>Europe:</b>	The field action is a Field Safety Corrective Action and is reported to the European Health Authorities.
<b>Australia/NZ:</b>	The affected product has been distributed to Australia/NZD. RPAC to assess if reporting is required locally.
<b>Brazil:</b>	The affected product has been distributed to Brazil. Biodina to assess if reporting is required locally.
<b>Japan:</b>	The affected product has been distributed to Japan. RKK to assess if reporting is required locally.

---

Prepared by: Mogens Thomasen  
Senior Product Manager  
Technical Product Support and Service