



PLEASE DELIVER UPON RECEIPT to LAB DIRECTOR or LAB MANAGER

Urgent Product Correction Notice

Our Ref: 5173-1 FSCA

Dear Valued bioMérieux Customer,

Our records indicate that your laboratory may be using MYLA® in conjunction with BCI CONNECT to send data to/from your LIS.

Description of Issue:

We have recently discovered that under certain conditions unwanted alterations to results could be applied when using the BCI Connect scripting feature.
This is a software anomaly.

As of today, we have one record where some AST results from VITEK® 2 were unexpectedly modified for one customer.

The Customer observed that drugs with result 'S' (sensitive) or 'IE' (insufficient evidence) in VITEK® 2 instrument and MYLA were changed to 'R' (resistant) in LIS.

For this customer there was no rule in place to change the value from 'S' to 'R' or 'IE' to 'R'.

However, there was a rule to change the value '(-)' into 'R' in field a4 (Final Category) in order to help the customer in the interpretation of results.

Due to the anomaly in the script's configuration, when the ExpertCategory value was missing in the result, the drug was disregarded in the counting of the index/number.

Then, during the 2nd step of the script, when the change in value was applied on the drug using its number/index, the count was exhaustive.


This resulted in a shift of index between a drug with "(-)" value and a drug where the "R" value was applied, which caused a value of "R" to be applied to the drug just before in the list, hence changing the result for this drug in the LIS from "S" to "R".



This issue can happen with MYLA V4.8.0/4.8.1 and MYLA V4.7.0/V4.7.1 in conjunction with BCI Connect.

In order to check if you are possibly concerned by the issue, you need to first confirm which configuration system you are using.

1. Log into the MYLA System Application.
2. On the main view, press the Supervision icon and select the BCI icon.
3. On the Systems view, on the right side, the BCI connect system is displayed.



If you are not using BCI Connect with MYLA and do not plan on using it in a near future, you are **NOT** concerned by the issue. Please fill the acknowledgement form indicating you are not using a BCI connection.

If you are using BCI Connect, you need to meet these criteria as well:

Please don't hesitate to contact your bioMérieux representative should you need assistance with this configuration verification.

- Use of MYLA V4.8.0/4.8.1 or V4.7.0/4.7.1 with BCI Connect
- AND**
- Use one of following LIS protocols: FTP, Shared Folder, RS232 (TCP or Serial)
- AND**
- Use the Copy/Move/Swap feature of the Scripting configuration
- AND**
- Use at least one of the following functions: Copy, Move, Swap, Delete, Uppercase, Lowercase, SetValue, Truncate, Prefix, Suffix, RemoveChars
- AND**
- The target field that is modified is an antibiotic result

You **ARE** impacted by this FSCA and you **HAVE** to implement the following corrections:

- Update your MYLA version as soon as possible with the incoming new software version of MYLA 4.8.2 that is already available on the market since 22-APR-2021 by contacting your bioMérieux local contact.



- Stop using the rules that allows to change the value in field a4 (Final Category) when using MYLA with BCI Connect (MYLA V4.7.0, 4.7.1 and patched version, MYLA V4.8.0, 4.8.1).
 - Schedule an appointment with a bioMérieux Field Specialist Engineer in order to proceed with the script changing.

A bioMérieux representative will contact you to proceed with these activities

Impact to patient/user:

bioMérieux has identified a potential safety risk leading in the worst case to a false resistant erroneous test result associated with this event.

Actions:

Please take the following actions at this time:

- Confirm this letter has been distributed to, and reviewed by, all appropriate personnel within your organization.
- Please store this letter with your bioMérieux instrument documentation.
- Please apply the mandatory corrections as needed.
- Complete the Acknowledgement Form and return it to your local bioMérieux representative. It is important that you return the acknowledgement form to bioMérieux even if you determine that this urgent product correction notice does not impact your facility. Please indicate your configuration on the acknowledgement form.

bioMérieux is committed to providing our customers with the highest quality products, and we apologize for any inconvenience this may have caused in your laboratory. If you have any questions or concerns, please contact your local bioMérieux representative.

Thank you for your continued use of bioMérieux products,

bioMérieux



Attachment A: Acknowledgement Form.

URGENT FIELD SAFETY NOTICE

FSCA 5173-1 - MYLA V4.8.0/4.8.1 and MYLA V4.7.0/4.7.1 - Changing AST result from Sensitive to Resistant when sending data to Laboratory Information System

**TO BE RETURNED TO YOUR BIOMÉRIEUX CUSTOMER SERVICE AT THE FOLLOWING
FAX NUMBER : XXXXXXXX**

Name of the laboratory:

City:

Customer number:

MYLA version	MYLA Serial number	BCI Connect user (Yes or no)

- I acknowledge receipt of the bioMérieux letter regarding the “MYLA V4.8.0/4.8.1 and MYLA V4.7.0/4.7.1 - Changing AST result from Sensitive to Resistant when sending data to Laboratory Information System”
- I will implement the required actions as indicated in the Urgent Field Safety Notice.
- Have you encountered impact on patients’ results, or reports of illness or injury related to the identified issue ?
 - Yes No

It is important that you complete this Acknowledgement Form and return it to bioMérieux.

DATE

SIGNATURE :