

Field Safety Notice

Product name & article: Hanger-Bar HeliQ – part number 72754 and 72755

FSCA ID: FSCA-2021-1

Action: Product inspection & correction action

Date: 2021-10-18

To: Human Care's "CUSTOMER"

Products concerned

This field safety notice applies to Human Care's

- Hanger-Bar 2P 35cm article no 72754
 - Hanger-Bar 2P 45cm article no 72755
- manufactured 2020-01.

Problem Description

Unfortunately, we have noticed that Loctite may be missing on the stop-screw which holds the hanger-bar hook in place. This is the hook which connects to the distance strap. Although the risk is small for hanger-bar coming loose from the hook – in that case the user may fall – we want you to check that the stop-screw is in place and that Loctite has been used and remains in the product. Below is an instruction on how to inspect and what actions may be required.

Inspection and actions

1. Unscrew the M4 screw (fig. 1) and check the guide threads (see yellow arrow in fig. 2) and that Loctite has been applied to the M4 screw.

If the M4 screw is missing or the threads are of poor quality (worn or damaged), the product should not be used. Contact Human Care.



fig. 1



fig. 2

2. Apply Loctite 243 (fig. 3) on the M4 screw (fig. 4).



fig. 3



fig. 4

3. Screw on the M4 screw again with a torque of 1.5 Nm (fig. 5) until it stops. Wait a few minutes until Loctite has dried. Then check that the hook can be swung freely (fig. 6).



fig. 5



fig. 6

Feedback to Human Care

Human Care hereby requests all customers to confirm that they have received this notice and that the necessary action is taken, following the above instruction. A form is sent out along with this notice, for easy re-reporting to Human Care. This must be done at latest 2021-11-30.

Contact information:

quality@humancaregroup.com

+ 46 8 665 35 00