

Single Registration Number (SRN): N/A



Urgent Field Safety Notices
Urgent Product Correction
Immediate Action Required

Date Issued June 9, 2022

Product

Product Description	List Number	Lot Number	US/ EU UDI
Alinity c Hemoglobin A1c Reagent Kit (HbA1c)	08P4320	59561UQ04	(01)00380740135591(17)220704(10)59561UQ04
		59779UQ05	(01)00380740135591(17)220820(10)59779UQ05
		60229UQ07	(01)00380740135591(17)221008(10)60229UQ07
		60540UQ09	(01)00380740135591(17)221219(10)60540UQ09
		61011UQ10	(01)00380740135591(17)230120(10)61011UQ10
		61349UQ01	(01)00380740135591(17)230402(10)61349UQ01
		61658UQ01	(01)00380740135591(17)230426(10)61658UQ01
Alinity c Hemoglobin A1c (HbA1c)	08P4377	59801UQ04	(01)00380740174989(17)220704(10)59801UQ04
		60763UQ09	(01)00380740174989(17)221219(10)60763UQ09

Explanation

This letter is to inform you of a potential for falsely elevated Hemoglobin %A1c and Hemoglobin A1c patient results to be generated using the Alinity c Hemoglobin A1c assay when analyzing whole blood or hemolysate samples with poor sample integrity, due to micro-clots and particulate matter, associated with low Total Hemoglobin values. To address this issue, a revision to the Total Hemoglobin low linearity value for the Alinity Hemoglobin A1c whole blood and hemolysate applications has been implemented.

The SPECIMEN COLLECTION AND PREPARATION FOR ANALYSIS and Assay Procedure, Whole Blood Preparation and Hemolysate Preparation sections of the Alinity c Hemoglobin A1c assay insert provide appropriate sample preparation instructions to ensure an optimal sample is presented for analysis.

The following whole blood and hemolysate Total Hemoglobin assay parameter changes have been implemented to reduce the potential for falsely elevated Hemoglobin %A1c and Hemoglobin A1c results to occur:

- Low Linearity value for the Total Hemoglobin Whole Blood assay file (THbWB) was increased from 0.7046 µmol/L to 53.8278 µmol/L.
- Low Linearity value for the Total Hemoglobin Hemolysate assay file (THbH) was increased from 16.3614 µmol/L to 1250 µmol/L.

There is no change to the measuring interval of the calculated assays.

Impact on Patient Results

There is a potential for falsely elevated Hemoglobin %A1c and Hemoglobin A1c patient results.

Necessary Actions to be Taken by Customer

1. Immediately install the updated assay file versions listed below (as applicable for your laboratory), and then manually configure the Low Linearity value. The assay files can be obtained from www.corelaboratory.abbott

Whole Blood Application

Assay File	Assay No.	Version
THbWB	1105	5
%A1cWB	3075	7
A1cWB	3074	7

Hemolysate Application

Assay File	Assay No.	Version
THbH	1107	7
%A1cH	3077	6
A1cH	3076	7

2. For the whole blood application, manually configure the **THbWB** Low Linearity to **53.8278**.
For the hemolysate application, manually configure the **THbH** Low Linearity to **1250**.

For detailed information on editing assay parameters, refer to Edit result settings of assay parameters (c-series), in the Alinity ci-series Operations Manual, Section 2.

3. Please review this letter with your Medical Director or Laboratory Management and follow your laboratory protocol regarding the need for reviewing previously reported patient results.
4. Complete and return the Customer Reply Form.
5. Please retain this letter for your laboratory records.
6. If you have forwarded the products listed above to other laboratories, please inform them of this Product Correction and provide to them a copy of this letter.

Contact Information

We sincerely regret any inconvenience this may have caused your laboratory. If you or any of the health care providers you serve have questions regarding this information, U.S. Customers please contact Customer Service at 1-877-4ABBOTT (available 24 hours a day, 7 days a week). Customers outside the U.S., please contact your local area Customer Service.

Adverse reactions or quality problems experienced with the use of this product may be reported to the FDA's MedWatch Adverse Event Reporting program either online (<http://www.fda.gov/MedWatch/report.htm>), by mail (<http://www.fda.gov/MedWatch/getforms.htm>), by phone (1-800-332-1088), or by fax (1-800-FDA-0178).

If you have experienced any patient or user injury associated with this Field Action, please immediately report the event to your local area Customer Service.