

**To customers of the
Dräger ventilators
Infinity Acute Care System - Workstation Critical Care (Evita V500),
Infinity Acute Care System – Workstation Neonatal Care (Babylog VN500)
and Evita V300**

January 2022

Important Safety Notice!

Cyber Security Improvement Activity

The following products are affected:

Infinity Acute Care System - Workstation Critical Care (Evita V500) with SW2.60 and lower.
Infinity Acute Care System – Workstation Neonatal Care (Babylog VN500) with SW2.60 and lower.
Evita V300 with SW2.60 and lower.

Madam or Sir,

Medical devices are increasingly operated in networked environments. The exchange of information among medical devices, hospital networks and the internet allow solutions that improve the treatment of patients by healthcare providers and can make health care better. At the same time, the operation in networked environments increases the risk of potential cybersecurity threats. A cybersecurity attack could impact the safety and effectiveness of medical devices.

The ventilators of the product family Evita V500/V300 and Babylog VN500 have been in use in many countries worldwide since 2007. They are highly accepted, reliable ventilators with in total more than 124 thousand years of operation. To date, Draeger has not received any report or evidence of any case of a cyberattack. Only a handful of customers operate their ventilators in a networked environment that allows an exchange of information between ventilators and Dräger Service Connect Gateway. Most devices are connected via a serial Medibus/Medibus X interface, which is not vulnerable to cyberattacks.

Theoretically, devices that are not connected to a network can also be exposed to potential cybersecurity threats. Such an attack would, however, require a direct physical access to the device. Someone would have to gain unauthorized access to an ICU and would have to modify each ventilator individually, potentially impacting the ventilation therapy.

Operating organizations should generally and continuously evaluate access restrictions to their operational environment.

Drägerwerk AG & Co. KGaA
Moislinger Allee 53-55
23558 Lübeck, Deutschland
Postanschrift:
23542 Lübeck, Deutschland
Tel. +49 451 882-0
Fax +49 451 882-2080
info@draeger.com
www.draeger.com
UID-Nr. DE135082211

Bankverbindungen:
Commerzbank AG, Lübeck
IBAN: DE95 2304 0022 0014 6795 00
Swift-Code: COBA DE FF 230
Sparkasse zu Lübeck
IBAN: DE15 2305 0101 0001 0711 17
Swift-Code: NOLADE21SPL

Sitz der Gesellschaft: Lübeck
Handelsregister:
Amtsgericht Lübeck HRB 7903 HL
Komplementär: Drägerwerk Verwaltungs AG
Sitz der Gesellschaft: Lübeck
Handelsregister:
Amtsgericht Lübeck HRB 7395 HL

Vorsitzender des Aufsichtsrats der
Drägerwerk AG & Co. KGaA und
Drägerwerk Verwaltungs AG:
Stefan Lauer
Vorstand:
Stefan Dräger (Vorsitzender)
Rainer Klug
Gert-Hartwig Lescow
Dr. Reiner Piske
Anton Schrofner

The abovementioned ventilators use specially hardened operating systems. However, one of the used operating systems cannot be updated any longer, and its vulnerabilities cannot be fixed anymore. Therefore, they are not prepared against potential cyber security threats to the same extent as newer devices. This includes those carried out with physical access. Draeger therefore recommends the following:

- Follow the recommendations in the instructions for use:
 - Limit or control physical access to the abovementioned ventilators.
 - Do not connect any non-approved devices to the USB, LAN and DVI interfaces.
 - Be attentive to notifications, alarms and alerts
- Consider closing/covering all unused USB, LAN and DVI interfaces

Should you decide to close/cover unused interfaces, Draeger offers to provide tools upon request to cover or close these data interfaces of the ventilators free of charge. If applicable, please contact your local Dräger organization. For the intended and authorized use of the interfaces the USB locks and interface covers can be removed with appropriate keys or tools.

Medibus and MedibusX as serial point-to-point communication protocols without network capabilities are not affected and may therefore safely be used. Should you operate the abovementioned devices within a network for remote service, please contact your local Dräger organization.


Please ensure that all users of the above-mentioned products and other concerned persons within your organization are made aware of this Important Safety Notice. If you have provided the products to third parties, please forward a copy of this information. The responsible authorities have been notified of this action. We apologize for any inconvenience but believe this to be an essential preventive measure to increase patient safety. We thank you for your continued support.

Frank Ralfs



Senior Product Manager
Business Area Respiratory Care
Business Unit Therapy
Drägerwerk AG & Co. KGaA

Sonja Hillmer



Director Post Market Surveillance
Quality & Regulatory Affairs
Medical Division
Drägerwerk AG & Co. KGaA