

Date: February 25, 2021

URGENT FIELD SAFETY NOTICE
Damaged Packaging for 37 lots of MBT & ATTUNE® Tibial Bases (S+)
Medical Device Product Recall (Removal) – Ref. 1932449

Product Subject to this Notice:

| Part Code | Lot / Batch | GTIN | Description |
|-------------|-------------|-------------|-------------|
| See Table 1 | See Table 1 | See Table 1 | See Table 1 |

PLEASE DISTRIBUTE THIS INFORMATION TO APPROPRIATE PERSONNEL AT YOUR FACILITY WHO MAY USE THE AFFECTED PRODUCT

Dear Valued Customer,

Please be advised that DePuy (Ireland) UC is initiating a voluntary removal for the 37 lots listed in Table 1 for MBT & ATTUNE (S+) Tibial Bases due to potential damaged packaging. These implant components are utilized in Total Knee Arthroplasty. The affected tibial bases are designed to be used in cemented applications.

A total of nine (9) complaints received to date reported packaging damage. See Image 1. Our investigation of these complaints determined that product packaging may be compromised, which may present a potential for sterile barrier integrity failure, resulting in the possibility of non-sterile implant or residual packaging materials on the tibial implants. See Image 2. To date, customer complaints have reported only surgical delay.

Image 1: Example of damaged packaging



Image 2: Example of residual material on implant surface



Potential Patient Impact

Damaged packaging could impact the seal and result in a sterility breach. Implantation of potentially contaminated product could lead to infection and pain. If presence of polyurethane residue is not detected intraoperatively, and is left in patient, this may potentially impair range of motion in a rotating platform construct; or, cause an inflammatory reaction. No adverse events or injuries have been reported for these issues. Surgical delay is the only harm that has been reported to date.

Patient Communication:

Physicians who have treated patients using the affected lots (see Table 1) should continue to follow up on those patients post-operatively according to the physician's standard of care.

Requested Actions:

Our records show that your facility received one or more of the affected lots manufactured on July 8, 2020. Please take the following actions:

1. Carefully review the information contained in this notice.
2. Return the affected products using the normal returns process. Work with your sales consultant to ensure affected devices are returned and to arrange replacement. To receive replacement product or reimbursement, customers must return the products subject to this notice.

3. Complete **all** fields of the attached Business Response Form. Please make sure to include your facility name and address, account number, name of person completing the form, title, email address, telephone number and signature in the spaces provided.
4. Forward this notice to any personnel in your facility who need to be informed.
5. If any of the products subject to this notice have been forwarded to another facility, contact that facility and provide a copy of this notice to the relevant personnel.
6. Post a copy of this notice in a visible area for awareness of this notice.

Contact Information:

At DePuy Synthes, we aim to keep people well at every age and every stage of life. Our primary goal is patient safety and customer satisfaction by providing high quality products. We apologize for any inconvenience that this notice may cause and appreciate your cooperation with our request. If you have any questions, please contact your local DePuy Synthes Sales Consultant.

This Field Safety Notice has been reported to the local competent authority.

Thank you for your attention and cooperation.

Kimberly Long
Field Action Coordinator
Email: OneMD-Field-Actions@its.jnj.com
Telephone: 574-221-8156-

TABLE 1: Products Subject to this Field Safety Notice (Removal)

| Product Code | Lot / Batch | GTIN | Description |
|--------------|-------------|----------------|-----------------------------|
| 129433130 | 9543538 | 10603295025788 | MBT CEM KEEL TIB TRAY SZ3 |
| 129433130 | 9553989 | 10603295025788 | MBT CEM KEEL TIB TRAY SZ3 |
| 129433130 | 9554547 | 10603295025788 | MBT CEM KEEL TIB TRAY SZ3 |
| 129433130 | 9554548 | 10603295025788 | MBT CEM KEEL TIB TRAY SZ3 |
| 129433130 | 9554551 | 10603295025788 | MBT CEM KEEL TIB TRAY SZ3 |
| 129433140 | 9548774 | 10603295025795 | MBT CEM KEEL TIB TRAY SZ4 |
| 129433140 | 9548777 | 10603295025795 | MBT CEM KEEL TIB TRAY SZ4 |
| 129433140 | 9554569 | 10603295025795 | MBT CEM KEEL TIB TRAY SZ4 |
| 150670003 | 9558300 | 10603295492054 | ATTUNE FB TIB BASE SZ 3 CEM |
| 150670004 | 9555958 | 10603295492061 | ATTUNE FB TIB BASE SZ 4 CEM |
| 150670005 | 9554006 | 10603295492078 | ATTUNE FB TIB BASE SZ 5 CEM |
| 150670005 | 9556397 | 10603295492078 | ATTUNE FB TIB BASE SZ 5 CEM |
| 150670006 | 9558618 | 10603295492085 | ATTUNE FB TIB BASE SZ 6 CEM |
| 150670007 | 9557992 | 10603295492092 | ATTUNE FB TIB BASE SZ 7 CEM |
| 150670008 | 9558238 | 10603295492108 | ATTUNE FB TIB BASE SZ 8 CEM |
| 150670008 | 9558239 | 10603295492108 | ATTUNE FB TIB BASE SZ 8 CEM |
| 150670009 | 9555613 | 10603295492115 | ATTUNE FB TIB BASE SZ 9 CEM |
| 150680003 | 9557372 | 10603295492153 | ATTUNE RP TIB BASE SZ 3 CEM |
| 150680003 | 9557439 | 10603295492153 | ATTUNE RP TIB BASE SZ 3 CEM |
| 150680003 | 9557440 | 10603295492153 | ATTUNE RP TIB BASE SZ 3 CEM |
| 150680003 | 9557450 | 10603295492153 | ATTUNE RP TIB BASE SZ 3 CEM |
| 150680003 | 9557460 | 10603295492153 | ATTUNE RP TIB BASE SZ 3 CEM |
| 150680004 | 9557444 | 10603295492160 | ATTUNE RP TIB BASE SZ 4 CEM |
| 150680004 | 9557445 | 10603295492160 | ATTUNE RP TIB BASE SZ 4 CEM |
| 150680004 | 9557446 | 10603295492160 | ATTUNE RP TIB BASE SZ 4 CEM |
| 150680004 | 9557453 | 10603295492160 | ATTUNE RP TIB BASE SZ 4 CEM |
| 150680004 | 9557455 | 10603295492160 | ATTUNE RP TIB BASE SZ 4 CEM |
| 150680005 | 9557376 | 10603295492177 | ATTUNE RP TIB BASE SZ 5 CEM |
| 150680005 | 9557377 | 10603295492177 | ATTUNE RP TIB BASE SZ 5 CEM |
| 150680005 | 9557457 | 10603295492177 | ATTUNE RP TIB BASE SZ 5 CEM |
| 150680005 | 9557471 | 10603295492177 | ATTUNE RP TIB BASE SZ 5 CEM |
| 150680005 | 9557473 | 10603295492177 | ATTUNE RP TIB BASE SZ 5 CEM |
| 150680005 | 9557476 | 10603295492177 | ATTUNE RP TIB BASE SZ 5 CEM |
| 150680006 | 9556100 | 10603295492184 | ATTUNE RP TIB BASE SZ 6 CEM |
| 150680006 | 9556127 | 10603295492184 | ATTUNE RP TIB BASE SZ 6 CEM |
| 150680006 | 9556128 | 10603295492184 | ATTUNE RP TIB BASE SZ 6 CEM |
| 150680007 | 9557434 | 10603295492191 | ATTUNE RP TIB BASE SZ 7 CEM |

Note: Unique Device Identifier (UDI): UDI = DI + PI

DI = Device Identifier = GTIN | PI = Production Identifier = Lot Number

URGENT: FIELD SAFETY NOTICE
Damaged Packaging for 37 lots of MBT & ATTUNE® Tibial Bases (S+)
Medical Device Product Recall (Removal) – Ref. 1932449

| Part Number | Part Description | GTIN | Lot Number |
|-------------|------------------|-------------|-------------|
| See Table 1 | See Table 1 | See Table 1 | See Table 1 |

Please complete this BRF **within 3 business days upon receipt of the notification** and email this form to OneMD-Field-Actions@its.jnj.com.

- The facility located the affected product in stock and is returning products (see below). A copy of this letter is being retained in the facility's records. I have read and understand the notification.

For product returns: Please call customer service following the typical returns process in order to acquire a return number prior to shipping product. Please enclose a photocopy of the completed Business Response Form as a packing slip in the box containing the product(s) you are returning. Return all identified affected product to: GMED Healthcare | JDE 8.12 Returns Dept. | ATTN: RETURNS FA 1932449 (SS NR-0156048) | Rue de Luxembourg 5 | ZI Trazegnies | BE - 6180 Courcelles | Belgium | TEL: 32-7-146-9404

- The facility has no affected product for return. A copy of this letter is being retained in the facility's records. I have read and understand the notification.

| | |
|---|-------------------------|
| Your Name: | Facility/Business Name: |
| Signed*: | Date: |
| Facility/Business Address, City: | |
| Account Number: | |
| J&J Sales Rep (as applicable): | |
| Date the notification was received: | |
| Email Address: | Telephone Number: |
| <i>*Your signature provides confirmation that you have received and understood this notification.</i> | |
| <i>Your comments are always welcome:</i> | |

Note: Unique Device Identifier (UDI): UDI = DI + PI
 DI = Device Identifier = GTIN | PI = Production Identifier = Lot Number

Please complete and return this page to your local DePuy Synthes sales organization.