

To all user of the following Artis zee systems

Product/Trade Name:	Artis zee ceiling, Artis zee III ceiling	E-mail	advancedtherapies-fsca.team@siemens-healthineers.com
Material number:	10094137, 10502502	Date	March, 2021
		Corrective Action ID	AX008/20/S; AX011/20/S

## **Customer Safety Information (CSI) for Field Safety Corrective Action**

**Subject: Abrasion of the cabling of Artis zee ceiling systems.**

Dear Customer,

We would like to inform you about a potential issue affecting the cabling of your Artis zee ceiling system and a corrective action that will be performed.

### **What is the issue and when does it occur?**

Some Artis zee ceiling systems show an increased abrasion of the cabling at the cable outlet of the inner C-arm. This issue occurs sporadically and is not considered as a systematic issue.

### **What is the impact on the operation of the system and what are the possible risks?**

Increased abrasion of the cabling may lead to damaged cabling which may result in limited functionality of the Artis zee ceiling system up to system failure. In this case, it might be necessary to stop clinical treatment or to continue treatment on an alternative system.

### **How was the issue identified and what is the root cause?**

The issue was detected during regular field observation and a subsequent inspection of potentially affected systems. The root cause of the issue is a suboptimal surface design in combination with an unfavorable routing of the cables in the area of the cable outlet of the inner C-arm.

#### **Siemens Healthcare GmbH**

Management: Bernhard Montag, President and Chief Executive Officer;  
Darleen Caron, Jochen Schmitz, Christoph Zindel

Chairman of the Supervisory Board: Ralf P. Thomas  
Registered office: Munich, Germany; Commercial Registry: Munich, HRB 213821  
WEEE-Reg.-No. DE 64872105  
SCF V12

**Which steps have to be taken by the user to avoid the possible risks associated with this issue?**

We strongly recommend to establish appropriate emergency procedures until the corrective action has been performed. In any case, please make sure that patient treatment can be continued in other ways if there is any possible danger for the safety of the patient.

**What actions are being taken by the manufacturer to mitigate possible risks?**

Our service organization will inspect the potentially affected area and will install an additional cover to improve the cable guidance and prevent further cable harness damage.

However, if damage caused by abrasion is too severe the cable harness will be replaced. The replacement will be addressed via additional update AX011/20/S.

**What is the efficiency of the corrective actions?**

The corrective action will mitigate the occurrence of the non-conformity.

**How will the corrective action be implemented?**

Our service organization will get in contact with you for an appointment to perform the corrective action. Please feel free to contact our service organization for an earlier appointment.

This letter will be distributed to affected customers as update AX009/20/S.

**What risks are there for patients who have previously been examined or treated using this system?**

There are no risks for patients who have previously been examined or treated.

Please ensure that all users of the affected products within your organization and others who may need to be informed will receive the safety relevant information provided with this notice and will comply with the recommendations therein.

We appreciate your understanding and cooperation with this safety advisory and ask you to immediately instruct your personnel accordingly. Please ensure that this safety advisory is retained in your product related records appropriately. Please keep this information at least until the measures have been finalized.

Please forward this safety information to any other organizations that could be affected by this measure.

If the device has been sold and is therefore no longer in your possession, please forward this safety notice to the new owner. We would also request you to inform us of the identity of the device's new owner where possible.

With best regards,

Siemens Healthcare GmbH  
Business Area Advanced Therapies (AT)

*i.V. Reinmar Killmann*  
Electronically signed by: Reinmar Killmann  
Reason: I am approving this document  
Date: Mar 5, 2021 09:58 GMT+1

Dr. Reinmar Killmann  
Vice President Project & Portfolio Management

*i.V. Johann Böck*  
Electronically signed by: Johann Boeck  
Reason: I am approving this document  
Date: Mar 5, 2021 09:28 GMT+1

Johann Böck  
Safety Officer Medical Devices AT